

Mass Laboratory

Calibration or Measurement Service			Measurand Level or Range			Measurement Conditions/Independent Variable		Expanded Uncertainty					Reference Standard used in calibration		Service Identifier
Quantity/Class	Instrument or Artifact	Instrument Type or Method	Minimum value	Maximum value	Units	Parameter	Specifications	Value	Units	Coverage Factor	Level of Confidence	Is the expanded uncertainty a relative one?	Standard	Source of traceability	
Mass	Non-automatic weighing instruments	Euramet/cg-18/v.04 SIM MWG7:2009 (identical to the original Euramet/cg18/v.02)		12	kg	Temperature		2·10 <sup>-6</sup>		2	95%	SI	E2	INM	INM-L01-28
						Relative Humidity									
Mass	Non-automatic weighing instruments	Euramet/cg-18/v.04 SIM MWG7:2009 (identical to the original Euramet/cg18/v.02)		150	kg	Temperature		2·10 <sup>-5</sup>		2	95%	SI	F2	INM	INM-L01-29
						Relative Humidity									

Last updated: 2020-02-24